

# Resource Summary Report

Generated by [NIF](#) on Sep 1, 2026

## Waters: 2695 HPLC System

RRID:SCR\_018649

Type: Tool

---

### Proper Citation

Waters: 2695 HPLC System (RRID:SCR\_018649)

---

### Resource Information

**URL:** <https://www.labx.com/product/waters-2695>

**Proper Citation:** Waters: 2695 HPLC System (RRID:SCR\_018649)

**Description:** High performance liquid chromatographic system with quaternary, low pressure mixing pump and inline vacuum degassing. It has wavelengths of 190 to 800 nm and flow rates from 50 uL per min to 5 mL per min with 2.1 mm ID columns and larger. Autosampler has maximum capacity of 120 vials (12x32, 2 mL) with programmable temperature control from 4 to 40C. Heated column compartment provides temperatures from 5 degrees above ambient to 65C.

**Synonyms:** Waters Alliance 2695 Separations Module

**Resource Type:** instrument resource

**Keywords:** HPLC System, Instrument, Equipment, USEDit, Waters, ABRF

**Availability:** Commercially available

**Resource Name:** Waters: 2695 HPLC System

**Resource ID:** SCR\_018649

**Alternate IDs:** Model\_Number\_2695, SCR\_020877

**Alternate URLs:** <https://www.waters.com/webassets/cms/library/docs/720004547en.pdf>

**Record Creation Time:** 20220129T080341+0000

**Record Last Update:** 20260829T182549+0000

---

### Ratings and Alerts

No rating or validation information has been found for Waters: 2695 HPLC System.

No alerts have been found for Waters: 2695 HPLC System.

---

## Data and Source Information

**Source:** [SciCrunch Registry](#)

---

## Usage and Citation Metrics

We found 1 mentions in open access literature.

**Listed below are recent publications.** The full list is available at [NIF](#) .

Badja C, et al. (2024) Insights from multi-omic modeling of neurodegeneration in xeroderma pigmentosum using an induced pluripotent stem cell system. Cell reports, 43(6), 114243.