

Resource Summary Report

Generated by [NIF](#) on Aug 27, 2026

Waters: 2695 HPLC System

RRID:SCR_018649

Type: Tool

Proper Citation

Waters: 2695 HPLC System (RRID:SCR_018649)

Resource Information

URL: <https://www.labx.com/product/waters-2695>

Proper Citation: Waters: 2695 HPLC System (RRID:SCR_018649)

Description: High performance liquid chromatographic system with quaternary, low pressure mixing pump and inline vacuum degassing. It has wavelengths of 190 to 800 nm and flow rates from 50 uL per min to 5 mL per min with 2.1 mm ID columns and larger. Autosampler has maximum capacity of 120 vials (12x32, 2 mL) with programmable temperature control from 4 to 40C. Heated column compartment provides temperatures from 5 degrees above ambient to 65C.

Synonyms: Waters Alliance 2695 Separations Module

Resource Type: instrument resource

Keywords: HPLC System, Instrument, Equipment, USEDit, Waters, ABRF

Availability: Commercially available

Resource Name: Waters: 2695 HPLC System

Resource ID: SCR_018649

Alternate IDs: Model_Number_2695, SCR_020877

Alternate URLs: <https://www.waters.com/webassets/cms/library/docs/720004547en.pdf>

Record Creation Time: 20220129T080341+0000

Record Last Update: 20260821T194132+0000

Ratings and Alerts

No rating or validation information has been found for Waters: 2695 HPLC System.

No alerts have been found for Waters: 2695 HPLC System.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We found 1 mentions in open access literature.

Listed below are recent publications. The full list is available at [NIF](#) .

Badja C, et al. (2024) Insights from multi-omic modeling of neurodegeneration in xeroderma pigmentosum using an induced pluripotent stem cell system. Cell reports, 43(6), 114243.